



Ref. Certif. No.

DE 3 - 40790

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE**
CERTIFICAT D'ESSAI OC

Product	General Medical Devices Vaporizer System		
Name and address of the applicant	STORZ & BICKEL GMBH & CO. KG In Grubenäcker 5-9 78532 Tuttlingen, GERMANY		
Name and address of the manufacturer	STORZ & BICKEL GMBH & CO. KG In Grubenäcker 5-9, 78532 Tuttlingen, GERMANY		
Name and address of the factory	STORZ & BICKEL GMBH & CO. KG In Grubenäcker 5-9, 78532 Tuttlingen, GERMANY		
Ratings and principal characteristics	Rated Input Voltage:	110- 120 V	220-240V
	Rated Frequency:	50-60Hz	
	Rated Input Power:	120VA	
	Protection Class:	II	
Trade mark (if any)	vapormed		
Model/type Ref.	Volcano Medic		
Additional information (if necessary)	Certificate DE 3 - 40082 issued 2014-04-14 is replaced by this version due to technical changes.		
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1-6:2010 IEC 60601-1-6:2010/AMD1:2013 IEC 60601-1-11:2010		
as shown in the Test Report Ref. No. which forms part of this certificate	713110770		

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2017-07-19
CB 17 07 63179 061

Stefan Reis



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service