

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	General Medical Devices Vaporizer
Name and address of the applicant	STORZ & BICKEL GMBH In Grubenäcker 5-9 78532 Tuttlingen GERMANY
Name and address of the manufacturer	STORZ & BICKEL GMBH In Grubenäcker 5-9, 78532 Tuttlingen, GERMANY
Name and address of the factory	STORZ & BICKEL GMBH In Grubenäcker 5-9, 78532 Tuttlingen, GERMANY
Ratings and principal characteristics	Rated input voltage: 5 - 15 VDC Rated input power: 10- 45 W Protection class: internally powered; II Degree of protection against electric shock: BF Degree of water protection: IP22 Intermittent operation: 7 min. ON / 30 min. OFF
Model/type Ref.	Mighty + Medic
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1-6:2010 IEC 60601-1-6:2010/AMD1:2013 IEC 60601-1-11:2015
as shown in the Test Report Ref. No. which forms part of this certificate	713181689

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 063179 0078 Rev. 00

Date, 2021-03-29



(Paul Bielawski)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service