



America

CERTIFICATE

No. U8 17 07 63179 064

Holder of Certificate: **STORZ & BICKEL GMBH & CO. KG**

In Grubenäcker 5-9
78532 Tuttlingen
GERMANY

Production Facility(ies): 63179

Certification Mark:



C US

Product: **General Medical Devices Vaporizer**

Model(s): **Mighty Medic**

Parameters:

Rated Input Voltage:	12 DC
Rated Input Power:	36 W
Protection Class:	internally powered
Degree of protection: against electric shock:	BF
Degree of waterprotection:	IP22

Remarks:
The certificate is valid for USA and Canada.
Other countries might have other requirements,
which were not part of this certification.

Tested according to: ANSI/AAMI ES 60601-1:2005/(R)2012
CAN/CSA-C22.2 No. 60601-1:14

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 713112373

Date, 2017-07-17

Page 1 of 1

